

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW110821\  
 Data File : VW023285.D  
 Acq On : 08 Nov 2021 17:44  
 Operator : SY/MD  
 Sample : M4522-15  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 BG330

Quant Time: Nov 09 05:17:26 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR110421WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Tue Nov 09 05:13:22 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 5.619  | 114   | 129291   | 5.000    | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117   | 131176   | 5.000    | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 60157    | 5.000    | ug/L   | 0.00     |
|                               |        |       |          |          |        |          |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 1.307  | 65    | 31893    | 3.938    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =      | 78.800%  |
| 7) Chloroethane-d5            | 1.568  | 69    | 29349    | 4.446    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =      | 89.000%  |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63    | 47842    | 3.155    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =      | 63.200%  |
| 20) 2-Butanone-d5             | 3.886  | 46    | 83086    | 59.542   | ug/L   | -0.01    |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =      | 119.080% |
| 24) Chloroform-d              | 4.352  | 84    | 78446    | 4.545    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 90.800%  |
| 26) 1,2-Dichloroethane-d4     | 5.034  | 65    | 38019    | 4.898    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 98.000%  |
| 32) Benzene-d6                | 5.053  | 84    | 146670   | 4.358    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 87.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.072  | 67    | 47257    | 4.770    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =      | 95.400%  |
| 41) Toluene-d8                | 7.320  | 98    | 123999   | 3.932    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 78.600%  |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 16435    | 4.375    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =      | 87.400%  |
| 46) 2-Hexanone-d5             | 8.088  | 63    | 62684    | 45.349   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =      | 90.700%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 31929    | 4.481    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =      | 89.600%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152   | 48817    | 4.873    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =      | 97.400%  |
|                               |        |       |          |          |        |          |
| Target Compounds              |        |       |          |          | Qvalue |          |
| 13) Acetone                   | 2.179  | 43    | 1852     | 2.172    | ug/L   | 98       |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV110821\  
Data File : VV023285.D  
Acq On : 08 Nov 2021 17:44  
Operator : SY/MD  
Sample : M4522-15  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BG330

Quant Time: Nov 09 05:17:26 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR110421WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Tue Nov 09 05:13:22 2021  
Response via : Initial Calibration

