

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110920\
 Data File : VV019302.D
 Acq On : 09 Nov 2020 14:22
 Operator : SY/MD
 Sample : VV1109WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK318

Quant Time: Nov 10 07:09:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 10 06:58:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	269683	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	264152	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	128360	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80031	38.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.90%
7) Chloroethane-d5	1.58	69	71663	43.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.88%
11) 1,1-Dichloroethene-d2	2.12	63	129434	33.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.18%
21) 2-Butanone-d5	3.91	46	123403	107.53	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.53%
24) Chloroform-d	4.38	84	169125	43.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.84%
26) 1,2-Dichloroethane-d4	5.05	65	120880	44.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.08%
32) Benzene-d6	5.07	84	313230	42.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.88%
36) 1,2-Dichloropropane-d6	6.09	67	102691	47.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.30%
41) Toluene-d8	7.34	98	278908	41.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.10%
43) trans-1,3-Dichloropropene-	7.64	79	51533	41.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.58%
47) 2-Hexanone-d5	8.11	63	73548	86.99	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.99%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	126099	43.46	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.92%
66) 1,2-Dichlorobenzene-d4	11.65	152	119663	46.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110920\
Data File : VV019302.D
Acq On : 09 Nov 2020 14:22
Operator : SY/MD
Sample : VV1109WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK318

Quant Time: Nov 10 07:09:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 10 06:58:38 2020
Response via : Initial Calibration

