

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110920\
 Data File : VV019304.D
 Acq On : 09 Nov 2020 15:34
 Operator : SY/MD
 Sample : VV1109MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK320

Quant Time: Nov 10 07:17:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 10 06:58:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	249973	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	247802	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	121119	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	75799	39.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.58%
7) Chloroethane-d5	1.58	69	68424	45.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.52%
11) 1,1-Dichloroethene-d2	2.12	63	123998	34.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.44%
21) 2-Butanone-d5	3.91	46	119289	112.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.14%
24) Chloroform-d	4.37	84	162803	45.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.18%
26) 1,2-Dichloroethane-d4	5.06	65	116949	46.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.98%
32) Benzene-d6	5.07	84	304404	44.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.98%
36) 1,2-Dichloropropane-d6	6.09	67	99354	48.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.26%
41) Toluene-d8	7.34	98	268246	42.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.16%
43) trans-1,3-Dichloropropene-	7.64	79	49025	42.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.76%
47) 2-Hexanone-d5	8.11	63	73539	92.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.72%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	121745	44.72	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.44%
66) 1,2-Dichlorobenzene-d4	11.65	152	116338	48.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110920\
Data File : VV019304.D
Acq On : 09 Nov 2020 15:34
Operator : SY/MD
Sample : VV1109MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK320

Quant Time: Nov 10 07:17:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 10 06:58:38 2020
Response via : Initial Calibration

