

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110920\
 Data File : VV019305.D
 Acq On : 09 Nov 2020 15:57
 Operator : SY/MD
 Sample : VV1109MBL02
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK321

Quant Time: Nov 10 07:20:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 10 06:58:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	260245	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	255207	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	122190	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	75418	37.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.10%
7) Chloroethane-d5	1.58	69	69587	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.44%
11) 1,1-Dichloroethene-d2	2.12	63	124372	33.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.90%
21) 2-Butanone-d5	3.91	46	124466	112.39	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.39%
24) Chloroform-d	4.38	84	166611	44.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.66%
26) 1,2-Dichloroethane-d4	5.06	65	119501	45.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.26%
32) Benzene-d6	5.07	84	308813	43.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.64%
36) 1,2-Dichloropropane-d6	6.09	67	101539	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.52%
41) Toluene-d8	7.34	98	271652	41.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.76%
43) trans-1,3-Dichloropropene-	7.64	79	50349	42.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.52%
47) 2-Hexanone-d5	8.11	63	74370	91.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.05%
56) 1,1,2,2-Tetrachloroethane-	10.24	84	125156	44.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.28%
66) 1,2-Dichlorobenzene-d4	11.65	152	118371	48.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110920\
Data File : VV019305.D
Acq On : 09 Nov 2020 15:57
Operator : SY/MD
Sample : VV1109MBL02
Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK321

Quant Time: Nov 10 07:20:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 10 06:58:38 2020
Response via : Initial Calibration

