

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110921\
 Data File : VV023327.D
 Acq On : 10 Nov 2021 12:58
 Operator : SY/MD
 Sample : M4545-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4617

Quant Time: Nov 11 00:41:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 11 00:38:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	124869	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	125361	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	58059	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	39804	5.088	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.800%
7) Chloroethane-d5	1.568	69	31875	5.000	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.000%
11) 1,1-Dichloroethene-d2	2.111	63	70487	4.813	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.200%
20) 2-Butanone-d5	3.889	46	70582	52.373	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.740%
24) Chloroform-d	4.352	84	73296	4.397	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
26) 1,2-Dichloroethane-d4	5.037	65	35071	4.678	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.053	84	147737	4.593	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
36) 1,2-Dichloropropane-d6	6.072	67	44452	4.695	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.320	98	126495	4.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	7.629	79	15236	4.244	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.088	63	51771	39.192	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.380%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29286	4.301	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	47473	4.911	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%
Target Compounds						Qvalue
3) Chloromethane	1.240	50	1138	0.110	ug/L	84
10) 1,1,2-Trichloro-1,2,2-...	2.114	101	1355	0.173	ug/L #	63
12) 1,1-Dichloroethene	2.121	96	29058	3.902	ug/L	94
13) Acetone	2.175	43	2272	2.759	ug/L	96
17) Methyl tert-butyl Ether	2.773	73	2318	0.141	ug/L #	90
29) 1,1,1-Trichloroethane	4.616	97	6669	0.438	ug/L	97
34) Trichloroethene	5.918	95	9116	0.978	ug/L	97
47) Tetrachloroethene	7.976	164	73348	9.083	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110921\
Data File : VV023327.D
Acq On : 10 Nov 2021 12:58
Operator : SY/MD
Sample : M4545-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4617

Quant Time: Nov 11 00:41:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 11 00:38:57 2021
Response via : Initial Calibration

