

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110921\
 Data File : VW023343.D
 Acq On : 10 Nov 2021 19:37
 Operator : SY/MD
 Sample : M4535-22
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4566

Quant Time: Nov 11 00:45:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 11 00:38:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	143844	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	139348	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68783	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	37203	4.129	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.600%
7) Chloroethane-d5	1.568	69	29164	3.971	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.400%
11) 1,1-Dichloroethene-d2	2.108	63	51726	3.066	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.400%
20) 2-Butanone-d5	3.892	46	80492	51.847	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.700%
24) Chloroform-d	4.352	84	75073	3.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.200%
26) 1,2-Dichloroethane-d4	5.034	65	37241	4.312	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
32) Benzene-d6	5.050	84	150704	4.215	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
36) 1,2-Dichloropropane-d6	6.072	67	44933	4.269	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.317	98	131507	3.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.600%
43) trans-1,3-Dichloroprop...	7.629	79	16187	4.056	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.200%
46) 2-Hexanone-d5	8.092	63	57207	38.960	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.920%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28322	3.742	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	51079	4.460	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.200%
Target Compounds					Qvalue	
13) Acetone	2.179	43	6786	7.154	ug/L	95
14) Carbon disulfide	2.298	76	3017	0.093	ug/L #	93
33) Benzene	5.108	78	7068	0.181	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110921\
Data File : VV023343.D
Acq On : 10 Nov 2021 19:37
Operator : SY/MD
Sample : M4535-22
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4566

Quant Time: Nov 11 00:45:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 11 00:38:57 2021
Response via : Initial Calibration

