

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110922\
 Data File : VV028855.D
 Acq On : 09 Nov 2022 10:31
 Operator : SY/MD
 Sample : VV1109WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK289

Quant Time: Nov 10 00:44:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 10 00:43:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	386775	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	354548	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	156410	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	174676	47.641	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.280%
7) Chloroethane-d5	1.564	69	157241	50.167	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.340%
11) 1,1-Dichloroethene-d2	2.105	63	263263	37.746	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.500%
21) 2-Butanone-d5	3.867	46	241669	90.546	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.550%
24) Chloroform-d	4.339	84	280361	50.699	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	5.021	65	192271	52.572	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.140%
32) Benzene-d6	5.040	84	585452	52.895	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.800%
36) 1,2-Dichloropropane-d6	6.063	67	202678	53.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.680%
41) Toluene-d8	7.307	98	502204	51.613	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.220%
43) trans-1,3-Dichloroprop...	7.616	79	77569	53.150	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.300%
47) 2-Hexanone-d5	8.082	63	145258	84.590	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.590%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	242048	48.306	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.620%
66) 1,2-Dichlorobenzene-d4	11.615	152	160570	52.008	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.020%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110922\
Data File : VV028855.D
Acq On : 09 Nov 2022 10:31
Operator : SY/MD
Sample : VV1109WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK289

Quant Time: Nov 10 00:44:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 10 00:43:48 2022
Response via : Initial Calibration

