

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111020\
 Data File : VV019322.D
 Acq On : 10 Nov 2020 14:43
 Operator : SY/MD
 Sample : VIBLK93
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK93

Quant Time: Nov 11 03:52:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 11 02:14:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	239909	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	232723	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	108280	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100764	52.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.70%
7) Chloroethane-d5	1.58	69	83386	56.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	112.60%
11) 1,1-Dichloroethene-d2	2.12	63	150822	41.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.16%
21) 2-Butanone-d5	3.92	46	121251	103.84	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.84%
24) Chloroform-d	4.38	84	173695	48.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.96%
26) 1,2-Dichloroethane-d4	5.06	65	123965	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
32) Benzene-d6	5.07	84	330894	50.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.24%
36) 1,2-Dichloropropane-d6	6.09	67	106622	54.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.34%
41) Toluene-d8	7.34	98	289089	45.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.18%
43) trans-1,3-Dichloropropene-	7.64	79	52165	45.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.60%
47) 2-Hexanone-d5	8.11	63	70744	82.99	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.99%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	124280	46.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.66%
64) 1,2-Dichlorobenzene-d4	11.65	152	114552	50.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.70%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111020\
Data File : VV019322.D
Acq On : 10 Nov 2020 14:43
Operator : SY/MD
Sample : VIBLK93
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK93

Quant Time: Nov 11 03:52:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 11 02:14:49 2020
Response via : Initial Calibration

