

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111020\
 Data File : VV019324.D
 Acq On : 10 Nov 2020 15:30
 Operator : SY/MD
 Sample : VIBLK94
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK94

Quant Time: Nov 11 03:59:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 11 02:14:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	238423	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	237986	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	107758	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	103655	54.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	109.42%
7) Chloroethane-d5	1.58	69	85955	58.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	116.80%
11) 1,1-Dichloroethene-d2	2.12	63	153297	42.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.06%
21) 2-Butanone-d5	3.91	46	121219	104.46	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.46%
24) Chloroform-d	4.38	84	174994	49.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.30%
26) 1,2-Dichloroethane-d4	5.06	65	124706	50.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.56%
32) Benzene-d6	5.07	84	334500	49.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.10%
36) 1,2-Dichloropropane-d6	6.09	67	106924	53.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.24%
41) Toluene-d8	7.34	98	292397	45.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.18%
43) trans-1,3-Dichloropropene-	7.64	79	52565	45.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.26%
47) 2-Hexanone-d5	8.11	63	71325	81.82	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.82%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	125582	46.27	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.54%
64) 1,2-Dichlorobenzene-d4	11.65	152	117153	51.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.50%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111020\
Data File : VV019324.D
Acq On : 10 Nov 2020 15:30
Operator : SY/MD
Sample : VIBLK94
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK94

Quant Time: Nov 11 03:59:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 11 02:14:49 2020
Response via : Initial Calibration

