

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
 Data File : VV023355.D
 Acq On : 11 Nov 2021 01:00
 Operator : SY/MD
 Sample : M4558-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB876

Quant Time: Nov 11 03:55:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 11 03:34:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	135257	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	131933	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62313	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	24183	2.854	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	57.000%	
7) Chloroethane-d5	1.568	69	24214	3.506	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	70.200%	
11) 1,1-Dichloroethene-d2	2.111	63	37733	2.379	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	47.600%#	
20) 2-Butanone-d5	3.889	46	99549	68.193	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	136.380%#	
24) Chloroform-d	4.352	84	72319	4.005	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.000%	
26) 1,2-Dichloroethane-d4	5.034	65	36576	4.504	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.000%	
32) Benzene-d6	5.053	84	133491	3.943	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	78.800%	
36) 1,2-Dichloropropane-d6	6.072	67	44951	4.511	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.200%	
41) Toluene-d8	7.320	98	110783	3.492	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.800%#	
43) trans-1,3-Dichloroprop...	7.625	79	13699	3.625	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	72.600%	
46) 2-Hexanone-d5	8.088	63	72187	51.925	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.840%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32908	4.592	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	48264	4.652	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.000%	
Target Compounds						
13) Acetone	2.178	43	2906	3.258	ug/L	91
18) trans-1,2-Dichloroethene	2.767	96	3631	0.366	ug/L	94
22) cis-1,2-Dichloroethene	3.911	96	178799	18.738	ug/L #	93
34) Trichloroethene	5.915	95	1200496	122.421	ug/L	99
47) Tetrachloroethene	7.979	164	3041	0.358	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
Data File : VV023355.D
Acq On : 11 Nov 2021 01:00
Operator : SY/MD
Sample : M4558-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB876

Quant Time: Nov 11 03:55:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 11 03:34:54 2021
Response via : Initial Calibration

