

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111021\
 Data File : VW023367.D
 Acq On : 11 Nov 2021 05:41
 Operator : SY/MD
 Sample : M4542-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGGF6

Quant Time: Nov 11 06:19:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 11 03:34:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	137765	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	133096	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	65057	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	25199	2.920	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	58.400%	
7) Chloroethane-d5	1.568	69	24523	3.486	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	69.800%	
11) 1,1-Dichloroethene-d2	2.111	63	38927	2.409	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	48.200%#	
20) 2-Butanone-d5	3.886	46	94773	63.740	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	127.480%	
24) Chloroform-d	4.352	84	73217	3.981	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	79.600%	
26) 1,2-Dichloroethane-d4	5.037	65	37256	4.504	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.000%	
32) Benzene-d6	5.053	84	138504	4.056	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.200%	
36) 1,2-Dichloropropane-d6	6.069	67	43176	4.295	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.800%	
41) Toluene-d8	7.317	98	114732	3.585	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.800%	
43) trans-1,3-Dichloroprop...	7.625	79	15151	3.975	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.400%	
46) 2-Hexanone-d5	8.088	63	65239	46.517	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.040%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	30367	4.201	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	84.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	50325	4.646	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.000%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	8133	0.712	ug/L	99
13) Acetone	2.175	43	10876	11.971	ug/L	94
34) Trichloroethene	5.921	95	4068	0.411	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
Data File : VV023367.D
Acq On : 11 Nov 2021 05:41
Operator : SY/MD
Sample : M4542-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGGF6

Quant Time: Nov 11 06:19:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 11 03:34:54 2021
Response via : Initial Calibration

