

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111021\
 Data File : VW023381.D
 Acq On : 11 Nov 2021 12:37
 Operator : SY/MD
 Sample : M4558-19DL 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB881DL

Quant Time: Nov 12 00:29:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 11 08:19:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	115801	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	113658	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	52544	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	32217	4.441	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.800%
7) Chloroethane-d5	1.568	69	26490	4.480	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.600%
11) 1,1-Dichloroethene-d2	2.108	63	46174	3.400	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.000%
20) 2-Butanone-d5	3.905	46	51398	41.124	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.240%
24) Chloroform-d	4.352	84	64571	4.177	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
26) 1,2-Dichloroethane-d4	5.037	65	31933	4.593	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
32) Benzene-d6	5.056	84	133357	4.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.075	67	39696	4.624	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.320	98	118524	4.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	7.628	79	14256	4.380	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.600%
46) 2-Hexanone-d5	8.095	63	42201	35.236	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.480%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	24742	4.008	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	42305	4.835	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.800%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.510	84	4887	0.485	ug/L	95
22) cis-1,2-Dichloroethene	3.921	96	13372	1.637	ug/L	# 87
25) Chloroform	4.384	83	5354	0.350	ug/L	79
34) Trichloroethene	5.918	95	68063	8.057	ug/L	97
47) Tetrachloroethene	7.985	164	1049	0.143	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
Data File : VV023381.D
Acq On : 11 Nov 2021 12:37
Operator : SY/MD
Sample : M4558-19DL 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 68 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB881DL

Quant Time: Nov 12 00:29:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 11 08:19:32 2021
Response via : Initial Calibration

