

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
 Data File : VV023392.D
 Acq On : 11 Nov 2021 16:55
 Operator : SY/MD
 Sample : VV1110WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 79 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK249

Quant Time: Nov 12 02:06:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 12 02:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	119921	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	117855	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53348	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	34960	4.654	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.000%
7) Chloroethane-d5	1.568	69	28577	4.667	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.108	63	49604	3.527	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.600%
20) 2-Butanone-d5	3.905	46	59463	45.943	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.880%
24) Chloroform-d	4.349	84	67943	4.244	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
26) 1,2-Dichloroethane-d4	5.030	65	33917	4.711	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	5.050	84	138524	4.581	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.072	67	41136	4.621	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.317	98	119154	4.205	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	7.625	79	14437	4.277	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.600%
46) 2-Hexanone-d5	8.091	63	51344	41.344	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.680%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28212	4.407	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	45564	5.129	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
Data File : VV023392.D
Acq On : 11 Nov 2021 16:55
Operator : SY/MD
Sample : VV1110WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 79 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK249

Quant Time: Nov 12 02:06:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 12 02:02:21 2021
Response via : Initial Calibration

