

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111021\
 Data File : VW023397.D
 Acq On : 11 Nov 2021 18:59
 Operator : SY/MD
 Sample : M4558-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB877

Quant Time: Nov 12 02:07:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 12 02:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	129509	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	122893	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	56013	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	32085	3.955	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.000%
7) Chloroethane-d5	1.568	69	26879	4.065	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.200%
11) 1,1-Dichloroethene-d2	2.108	63	46032	3.031	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.600%
20) 2-Butanone-d5	3.896	46	75223	53.817	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.640%
24) Chloroform-d	4.349	84	67042	3.877	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
26) 1,2-Dichloroethane-d4	5.037	65	33432	4.300	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.000%
32) Benzene-d6	5.053	84	133720	4.241	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
36) 1,2-Dichloropropane-d6	6.072	67	40002	4.310	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.200%
41) Toluene-d8	7.317	98	114607	3.879	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.600%
43) trans-1,3-Dichloroprop...	7.629	79	14317	4.068	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.400%
46) 2-Hexanone-d5	8.092	63	54557	42.130	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.260%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	27429	4.109	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	43854	4.702	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.000%
Target Compounds						Qvalue
5) Vinyl chloride	1.310	62	1782	0.166	ug/L #	66
13) Acetone	2.185	43	2983	3.493	ug/L	87
17) Methyl tert-butyl Ether	2.770	73	2668	0.157	ug/L #	80
18) trans-1,2-Dichloroethene	2.764	96	943	0.099	ug/L	73
22) cis-1,2-Dichloroethene	3.912	96	44493	4.870	ug/L #	88
27) 1,2-Dichloroethane	5.140	62	9162	1.008	ug/L	99
33) Benzene	5.105	78	18124	0.528	ug/L	100
34) Trichloroethene	5.918	95	30647	3.355	ug/L	99
42) Toluene	7.397	91	4143	0.113	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
Data File : VV023397.D
Acq On : 11 Nov 2021 18:59
Operator : SY/MD
Sample : M4558-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB877

Quant Time: Nov 12 02:07:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 12 02:02:21 2021
Response via : Initial Calibration

