

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
 Data File : VV023351.D
 Acq On : 10 Nov 2021 23:26
 Operator : SY/MD
 Sample : VV1110WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK247

Quant Time: Nov 11 03:35:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 11 03:34:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	125235	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	126242	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	59659	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	37622	4.795	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.000%
7) Chloroethane-d5	1.568	69	31076	4.860	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.200%
11) 1,1-Dichloroethene-d2	2.111	63	53319	3.630	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.600%
20) 2-Butanone-d5	3.899	46	70505	52.163	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.320%
24) Chloroform-d	4.355	84	72736	4.350	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
26) 1,2-Dichloroethane-d4	5.040	65	36752	4.888	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.053	84	147377	4.550	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	6.072	67	43698	4.583	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.600%
41) Toluene-d8	7.320	98	129044	4.251	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	7.628	79	15243	4.216	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.400%
46) 2-Hexanone-d5	8.091	63	57555	43.266	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.540%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29893	4.359	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	49982	5.031	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111021\
Data File : VV023351.D
Acq On : 10 Nov 2021 23:26
Operator : SY/MD
Sample : VV1110WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK247

Quant Time: Nov 11 03:35:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 11 03:34:54 2021
Response via : Initial Calibration

