

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
 Data File : VV028868.D
 Acq On : 10 Nov 2022 14:43
 Operator : SY/MD
 Sample : N5574-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AA0

Quant Time: Nov 11 02:06:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 02:00:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	21862	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	19703	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	9613	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	7344	0.353	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	70.000%
4) 1,2-Dichloroethane-d4	5.042	65	6564	0.516	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.000%
7) 1,2-Dichloropropane-d6	6.068	67	8653	0.537	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.000%
8) Toluene-d8	7.317	98	21714	0.478	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.198	84	5049	0.485	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	96.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
Data File : VV028868.D
Acq On : 10 Nov 2022 14:43
Operator : SY/MD
Sample : N5574-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AA0

Quant Time: Nov 11 02:06:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 02:00:02 2022
Response via : Initial Calibration

