

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
 Data File : VV028862.D
 Acq On : 10 Nov 2022 11:04
 Operator : SY/MD
 Sample : VSTD0.516
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5216

Quant Time: Nov 11 01:56:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 01:55:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	21675	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	19584	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	9534	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	11275	0.547	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	6812	0.540	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	8835	0.552	ug/L	0.00
8) Toluene-d8	7.317	98	24452	0.542	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.198	84	5604	0.541	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	9694	0.505	ug/L	100
6) Trichloroethene	5.925	95	7261	0.496	ug/L	100
9) 1,2-Dibromoethane	8.353	107	4279	0.504	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	3809	0.510	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.409	75	553	0.435	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
Data File : VV028862.D
Acq On : 10 Nov 2022 11:04
Operator : SY/MD
Sample : VSTD0.516
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5216

Quant Time: Nov 11 01:56:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 01:55:41 2022
Response via : Initial Calibration

