

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
 Data File : VV028863.D
 Acq On : 10 Nov 2022 11:26
 Operator : SY/MD
 Sample : VSTD00117
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001217

Quant Time: Nov 11 01:56:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 01:55:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	20911	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	19367	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	9646	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	21468	1.080	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	13139	1.080	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	17038	1.075	ug/L	0.00
8) Toluene-d8	7.317	98	47326	1.060	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.198	84	10890	1.063	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	19673	1.063	ug/L	95
6) Trichloroethene	5.925	95	14650	1.013	ug/L	98
9) 1,2-Dibromoethane	8.353	107	8771	1.046	ug/L #	99
12) 1,2,3-Trichloropropane	10.267	75	7903	1.045	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.409	75	1271	0.988	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
Data File : VV028863.D
Acq On : 10 Nov 2022 11:26
Operator : SY/MD
Sample : VSTD00117
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001217

Quant Time: Nov 11 01:56:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 01:55:41 2022
Response via : Initial Calibration

