

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
 Data File : VV028867.D
 Acq On : 10 Nov 2022 14:02
 Operator : SY/MD
 Sample : N5568-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 X3D61

Quant Time: Nov 11 02:06:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 02:00:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	22317	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	19843	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	9464	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	10846	0.511	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	5.042	65	6491	0.500	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 100.000%	
7) 1,2-Dichloropropane-d6	6.068	67	8380	0.516	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 104.000%	
8) Toluene-d8	7.317	98	23074	0.505	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 100.000%	
10) 1,1,2,2-Tetrachloroeth...	10.198	84	5022	0.479	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 96.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	19472	0.986	ug/L	95
6) Trichloroethene	5.901	95	29839	2.013	ug/L	90
9) 1,2-Dibromoethane	8.353	107	5460	0.635	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	1550	0.209	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.409	75	285	0.226	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
Data File : VV028867.D
Acq On : 10 Nov 2022 14:02
Operator : SY/MD
Sample : N5568-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
X3D61

Quant Time: Nov 11 02:06:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 02:00:02 2022
Response via : Initial Calibration

