

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
 Data File : VV028869.D
 Acq On : 10 Nov 2022 15:08
 Operator : SY/MD
 Sample : N5574-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AA1

Quant Time: Nov 11 02:06:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 02:00:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	20268	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	18499	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	8780	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	10257	0.532	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 106.000%	
4) 1,2-Dichloroethane-d4	5.042	65	6202	0.526	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
7) 1,2-Dichloropropane-d6	6.068	67	8156	0.539	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.317	98	22449	0.527	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
10) 1,1,2,2-Tetrachloroeth...	10.198	84	4952	0.506	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 102.000%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
Data File : VV028869.D
Acq On : 10 Nov 2022 15:08
Operator : SY/MD
Sample : N5574-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AA1

Quant Time: Nov 11 02:06:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 02:00:02 2022
Response via : Initial Calibration

