

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
 Data File : VV028860.D
 Acq On : 10 Nov 2022 09:50
 Operator : SY/MD
 Sample : VSTD0.0514
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05214

Quant Time: Nov 11 01:56:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 01:55:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	23833	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	21509	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	10643	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	1004	0.044	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	596	0.043	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	770	0.044	ug/L	0.00
8) Toluene-d8	7.332	98	2262	0.046	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.198	84	506	0.044	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	956	0.045	ug/L #	60
6) Trichloroethene	5.925	95	796	0.050	ug/L	97
9) 1,2-Dibromoethane	8.368	107	438	0.047	ug/L #	89
12) 1,2,3-Trichloropropane	10.267	75	414	0.050	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.409	75	81	0.057	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111022\
Data File : VV028860.D
Acq On : 10 Nov 2022 09:50
Operator : SY/MD
Sample : VSTD0.0514
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05214

Quant Time: Nov 11 01:56:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 01:55:41 2022
Response via : Initial Calibration

