

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111023\
 Data File : VW032882.D
 Acq On : 10 Nov 2023 11:24
 Operator : SY/MD
 Sample : 05326-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E45K7

Quant Time: Nov 10 22:50:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	345495	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	380480	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	178842	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	137449	41.827	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.660%	
7) Chloroethane-d5	1.535	69	121425	44.181	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	88.360%	
11) 1,1-Dichloroethene-d2	2.063	65	62549	46.083	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	92.160%	
21) 2-Butanone-d5	3.786	46	128158	89.913	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	89.910%	
24) Chloroform-d	4.252	84	253801	48.387	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.780%	
26) 1,2-Dichloroethane-d4	4.944	65	165000	50.614	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.220%	
32) Benzene-d6	4.963	84	486011	49.681	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.360%	
36) 1,2-Dichloropropane-d6	5.992	67	160684	54.065	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	108.140%	
41) Toluene-d8	7.246	98	477634	49.585	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.180%	
43) trans-1,3-Dichloroprop...	7.555	79	73156	47.262	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	94.520%	
47) 2-Hexanone-d5	8.027	63	97271	89.650	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	89.650%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	214278	47.511	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	95.020%	
66) 1,2-Dichlorobenzene-d4	11.564	152	192707	55.915	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	111.840%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032882.D
Acq On : 10 Nov 2023 11:24
Operator : SY/MD
Sample : 05326-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E45K7

Quant Time: Nov 10 22:50:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

