

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111023\
 Data File : VW032883.D
 Acq On : 10 Nov 2023 11:47
 Operator : SY/MD
 Sample : 05333-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCGD5

Quant Time: Nov 10 22:50:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	317376	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	348183	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	156899	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	115625	38.303	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	76.600%	
7) Chloroethane-d5	1.532	69	103159	40.860	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	81.720%	
11) 1,1-Dichloroethene-d2	2.060	65	51253	41.106	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	82.220%	
21) 2-Butanone-d5	3.786	46	114929	87.775	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	87.780%	
24) Chloroform-d	4.252	84	208253	43.221	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.440%	
26) 1,2-Dichloroethane-d4	4.944	65	136344	45.530	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.060%	
32) Benzene-d6	4.963	84	401497	44.848	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.700%	
36) 1,2-Dichloropropane-d6	5.989	67	134296	49.378	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	98.760%	
41) Toluene-d8	7.246	98	379997	43.108	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.220%	
43) trans-1,3-Dichloroprop...	7.555	79	60006	42.363	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	84.720%	
47) 2-Hexanone-d5	8.027	63	77986	78.543	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	78.540%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	177299	42.959	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	85.920%	
66) 1,2-Dichlorobenzene-d4	11.564	152	156821	51.867	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.740%	
Target Compounds					Qvalue	
13) Acetone	2.114	43	17603	8.032	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032883.D
Acq On : 10 Nov 2023 11:47
Operator : SY/MD
Sample : 05333-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCGD5

Quant Time: Nov 10 22:50:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

