

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111023\
 Data File : VV032895.D
 Acq On : 10 Nov 2023 16:29
 Operator : SY/MD
 Sample : 05296-05 10X
 Misc : 4.61gm/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH3R6

Quant Time: Nov 10 22:52:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	412885	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	420518	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	200786	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	168958	43.024	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.040%
7) Chloroethane-d5	1.526	69	154948	47.177	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.360%
11) 1,1-Dichloroethene-d2	2.056	65	76861	47.385	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.760%
21) 2-Butanone-d5	3.786	46	165080	96.913	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.910%
24) Chloroform-d	4.249	84	319367	50.950	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.900%
26) 1,2-Dichloroethane-d4	4.940	65	211984	54.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.820%
32) Benzene-d6	4.960	84	661497	61.181	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	122.360%
36) 1,2-Dichloropropane-d6	5.989	67	200389	61.005	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	122.000%#
41) Toluene-d8	7.246	98	587855	55.217	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.440%
43) trans-1,3-Dichloroprop...	7.554	79	92849	54.274	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.540%
47) 2-Hexanone-d5	8.024	63	125160	104.371	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.370%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	264821	53.128	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.260%
66) 1,2-Dichlorobenzene-d4	11.564	152	242172	62.588	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	125.180%#
Target Compounds						
					Qvalue	
46) Tetrachloroethene	7.911	164	5086	1.816	ug/L	92
72) 1,2,3-Trichlorobenzene	13.683	180	15820	4.001	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032895.D
Acq On : 10 Nov 2023 16:29
Operator : SY/MD
Sample : 05296-05 10X
Misc : 4.61gm/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH3R6

Quant Time: Nov 10 22:52:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

