

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111023\
 Data File : VW032900.D
 Acq On : 10 Nov 2023 18:26
 Operator : SY/MD
 Sample : VIBLK243
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK243

Quant Time: Nov 10 22:53:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	312882	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	319862	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	139552	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	100434	33.749	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	67.500%		
7) Chloroethane-d5	1.532	69	100381	40.331	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	80.660%		
11) 1,1-Dichloroethene-d2	2.059	65	46054	37.467	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.940%		
21) 2-Butanone-d5	3.786	46	121941	94.468	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	94.470%		
24) Chloroform-d	4.252	84	220570	46.435	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.880%		
26) 1,2-Dichloroethane-d4	4.944	65	154977	52.495	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	105.000%		
32) Benzene-d6	4.963	84	441316	53.661	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	107.320%		
36) 1,2-Dichloropropane-d6	5.992	67	141279	56.545	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	113.080%		
41) Toluene-d8	7.246	98	362854	44.808	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	89.620%		
43) trans-1,3-Dichloroprop...	7.554	79	63675	48.933	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.860%		
47) 2-Hexanone-d5	8.027	63	87176	95.572	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	95.570%		
56) 1,1,2,2-Tetrachloroeth...	10.156	84	199433	52.600	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	105.200%		
66) 1,2-Dichlorobenzene-d4	11.564	152	149053	55.425	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	110.860%		
Target Compounds						Qvalue
46) Tetrachloroethene	7.908	164	1450	0.681	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032900.D
Acq On : 10 Nov 2023 18:26
Operator : SY/MD
Sample : VIBLK243
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK243

Quant Time: Nov 10 22:53:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

