

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111023\
 Data File : VW032913.D
 Acq On : 11 Nov 2023 00:16
 Operator : SY/MD
 Sample : 05296-15 10X
 Misc : 6.46gm/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH3T1

Quant Time: Nov 11 01:31:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	442188	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	430309	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	213053	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	131969	31.378	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	62.760%		
7) Chloroethane-d5	1.526	69	125624	35.714	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	71.420%		
11) 1,1-Dichloroethene-d2	2.056	65	61979	35.678	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	71.360%		
21) 2-Butanone-d5	3.786	46	135345	74.191	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	74.190%		
24) Chloroform-d	4.249	84	261590	38.967	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	77.940%		
26) 1,2-Dichloroethane-d4	4.941	65	170938	40.970	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	81.940%		
32) Benzene-d6	4.960	84	550638	49.769	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.540%		
36) 1,2-Dichloropropane-d6	5.989	67	166905	49.655	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.320%		
41) Toluene-d8	7.246	98	497947	45.708	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.420%		
43) trans-1,3-Dichloroprop...	7.555	79	77367	44.195	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.380%		
47) 2-Hexanone-d5	8.024	63	107190	87.352	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	87.350%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	220432	43.216	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	86.440%		
66) 1,2-Dichlorobenzene-d4	11.561	152	209230	50.961	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.920%		
Target Compounds						
46) Tetrachloroethene	7.912	164	2317	0.808	ug/L	98
53) m,p-Xylene	9.085	106	4101	0.752	ug/L	93
70) 1,2,4-trichlorobenzene	13.201	180	14054	3.344	ug/L	99
72) 1,2,3-Trichlorobenzene	13.683	180	22673	5.404	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032913.D
Acq On : 11 Nov 2023 00:16
Operator : SY/MD
Sample : 05296-15 10X
Misc : 6.46gm/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH3T1

Quant Time: Nov 11 01:31:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

