

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111023\
 Data File : VW032917.D
 Acq On : 11 Nov 2023 01:48
 Operator : SY/MD
 Sample : VIBLK244
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK244

Quant Time: Nov 11 03:44:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	413755	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	409886	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	195289	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	121711	30.928	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	61.860%
7) Chloroethane-d5	1.535	69	112437	34.161	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.320%#
11) 1,1-Dichloroethene-d2	2.063	65	57069	35.109	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.220%
21) 2-Butanone-d5	3.786	46	129354	75.780	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.780%
24) Chloroform-d	4.252	84	254774	40.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.120%
26) 1,2-Dichloroethane-d4	4.944	65	166892	42.749	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.500%
32) Benzene-d6	4.963	84	498305	47.283	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.560%
36) 1,2-Dichloropropane-d6	5.992	67	154539	48.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.540%
41) Toluene-d8	7.246	98	426844	41.133	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.260%
43) trans-1,3-Dichloroprop...	7.554	79	65387	39.213	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.420%
47) 2-Hexanone-d5	8.027	63	100516	85.994	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.990%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	209092	43.036	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.080%
66) 1,2-Dichlorobenzene-d4	11.564	152	179830	47.785	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032917.D
Acq On : 11 Nov 2023 01:48
Operator : SY/MD
Sample : VIBLK244
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK244

Quant Time: Nov 11 03:44:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

