

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
 Data File : VV032902.D
 Acq On : 10 Nov 2023 19:59
 Operator : SY/MD
 Sample : VV1110MBL02
 Misc : 5.00gm/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK216

Quant Time: Nov 10 20:40:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	369845	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	374963	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	176465	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	150945	42.910	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.820%
7) Chloroethane-d5	1.532	69	129876	44.145	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.280%
11) 1,1-Dichloroethene-d2	2.060	65	66014	45.434	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.860%
21) 2-Butanone-d5	3.786	46	142004	93.067	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.070%
24) Chloroform-d	4.253	84	264799	47.160	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.320%
26) 1,2-Dichloroethane-d4	4.944	65	171063	49.019	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.040%
32) Benzene-d6	4.963	84	521153	54.057	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.120%
36) 1,2-Dichloropropane-d6	5.989	67	168178	57.419	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	114.840%
41) Toluene-d8	7.246	98	491884	51.816	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.640%
43) trans-1,3-Dichloroprop...	7.555	79	76252	49.987	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.980%
47) 2-Hexanone-d5	8.027	63	104419	97.654	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.650%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	220712	49.658	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.320%
66) 1,2-Dichlorobenzene-d4	11.564	152	197087	57.957	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111023\
Data File : VV032902.D
Acq On : 10 Nov 2023 19:59
Operator : SY/MD
Sample : VV1110MBL02
Misc : 5.00gm/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK216

Quant Time: Nov 10 20:40:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

