

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
 Data File : VV028893.D
 Acq On : 11 Nov 2022 22:28
 Operator : SY/MD
 Sample : N5592-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AC0

Quant Time: Nov 12 02:58:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 22:27:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	307956	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	282192	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	135437	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	85914	4.576	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.600%	
7) Chloroethane-d5	1.568	69	79242	4.801	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.000%	
11) 1,1-Dichloroethene-d2	2.105	63	130354	3.810	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.200%	
20) 2-Butanone-d5	3.886	46	186669	59.396	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.800%	
24) Chloroform-d	4.342	84	190343	4.938	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.800%	
26) 1,2-Dichloroethane-d4	5.031	65	87681	5.056	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.200%	
32) Benzene-d6	5.047	84	404749	4.805	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
36) 1,2-Dichloropropane-d6	6.066	67	118598	5.042	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.800%	
41) Toluene-d8	7.310	98	355078	4.683	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.600%	
43) trans-1,3-Dichloroprop...	7.622	79	34328	4.475	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.400%	
46) 2-Hexanone-d5	8.088	63	164681	50.293	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.580%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	82169	5.039	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.800%	
66) 1,2-Dichlorobenzene-d4	11.615	152	119985	4.952	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.000%	
Target Compounds						
12) 1,1-Dichloroethene	2.117	96	19669	1.055	ug/L	1
19) 1,1-Dichloroethane	3.188	63	7894	0.223	ug/L	92
22) cis-1,2-Dichloroethene	3.912	96	20323	0.925	ug/L	94
34) Trichloroethene	5.912	95	150417	6.813	ug/L	98
47) Tetrachloroethene	7.969	164	474312	26.832	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028893.D
Acq On : 11 Nov 2022 22:28
Operator : SY/MD
Sample : N5592-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AC0

Quant Time: Nov 12 02:58:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 22:27:09 2022
Response via : Initial Calibration

