

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111122\
 Data File : VW028899.D
 Acq On : 12 Nov 2022 00:55
 Operator : SY/MD
 Sample : N5592-22MS
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AE1MS

Quant Time: Nov 12 03:00:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 22:27:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	306037	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	281903	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	135977	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	167221	8.963	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	179.200%#	
7) Chloroethane-d5	1.568	69	155899	9.505	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	190.000%#	
11) 1,1-Dichloroethene-d2	2.105	63	238474	7.014	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	140.200%#	
20) 2-Butanone-d5	3.879	46	387205	123.977	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	247.960%#	
24) Chloroform-d	4.342	84	381073	9.948	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	199.000%#	
26) 1,2-Dichloroethane-d4	5.027	65	178108	10.334	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	206.600%#	
32) Benzene-d6	5.043	84	817833	9.719	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	194.400%#	
36) 1,2-Dichloropropane-d6	6.066	67	237430	10.105	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	202.000%#	
41) Toluene-d8	7.310	98	728506	9.618	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	192.400%#	
43) trans-1,3-Dichloroprop...	7.619	79	72160	9.416	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	188.400%#	
46) 2-Hexanone-d5	8.085	63	352399	107.733	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	215.460%#	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	164164	10.077	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	201.600%#	
66) 1,2-Dichlorobenzene-d4	11.615	152	253459	10.419	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	208.400%#	
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.751	101	566630	17.400	ug/L	99
25) Chloroform	4.365	83	17013	0.462	ug/L	81
38) Bromodichloromethane	6.516	83	2642	0.105	ug/L #	96
47) Tetrachloroethene	7.973	164	25612	1.450	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028899.D
Acq On : 12 Nov 2022 00:55
Operator : SY/MD
Sample : N5592-22MS
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AE1MS

Quant Time: Nov 12 03:00:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 22:27:09 2022
Response via : Initial Calibration

