

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111122\
 Data File : VW028914.D
 Acq On : 12 Nov 2022 07:29
 Operator : SY/MD
 Sample : N5590-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBMY4

Quant Time: Nov 14 00:25:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 12 03:43:58 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards							
1) 1,4-Difluorobenzene		5.622	114	247628	5.000	ug/L	0.01
28) Chlorobenzene-d5		8.866	117	277804	5.000	ug/L	0.02
58) 1,4-Dichlorobenzene-d4		11.246	152	132247	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.301	65	77218	5.115	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.200%	
7) Chloroethane-d5		1.587	69	76822	5.788	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.800%	
11) 1,1-Dichloroethene-d2		2.153	63	98679	3.587	ug/L	0.04
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.800%	
20) 2-Butanone-d5		4.037	46	127282	50.367	ug/L	0.15
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.740%	
24) Chloroform-d		4.394	84	158292	5.107	ug/L	0.05
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.200%	
26) 1,2-Dichloroethane-d4		5.069	65	72953	5.231	ug/L	0.04
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%	
32) Benzene-d6		5.072	84	293565	3.540	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.800%	
36) 1,2-Dichloropropane-d6		6.088	67	121389	5.243	ug/L	0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.800%	
41) Toluene-d8		7.349	98	346917	4.648	ug/L	0.04
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%	
43) trans-1,3-Dichloroprop...		7.654	79	30292	4.011	ug/L	0.04
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%	
46) 2-Hexanone-d5		8.127	63	146918	45.577	ug/L	0.04
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.160%	
56) 1,1,2,2-Tetrachloroeth...		10.220	84	77970	4.857	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4		11.619	152	120697	5.101	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.000%	
Target Compounds							Qvalue
13) Acetone		2.310	43	134936	83.006	ug/L	88
16) Methylene chloride		2.545	84	29501	1.522	ug/L	91
22) cis-1,2-Dichloroethene		3.960	96	16815	0.951	ug/L	79
34) Trichloroethene		5.915	95	32345	1.488	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028914.D
Acq On : 12 Nov 2022 07:29
Operator : SY/MD
Sample : N5590-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBMY4

Quant Time: Nov 14 00:25:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 12 03:43:58 2022
Response via : Initial Calibration

