

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
 Data File : VV028917.D
 Acq On : 12 Nov 2022 08:43
 Operator : SY/MD
 Sample : N5602-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AG7

Quant Time: Nov 14 00:26:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 12 03:43:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	286443	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	277564	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	139872	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	73221	4.193	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%
7) Chloroethane-d5	1.571	69	71436	4.653	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.000%
11) 1,1-Dichloroethene-d2	2.114	63	103479	3.252	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.000%
20) 2-Butanone-d5	3.982	46	124694	42.656	ug/L	0.10
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.320%
24) Chloroform-d	4.349	84	157038	4.380	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
26) 1,2-Dichloroethane-d4	5.037	65	72205	4.476	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	5.043	84	305919	3.692	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.800%
36) 1,2-Dichloropropane-d6	6.069	67	77030	3.330	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	66.600%
41) Toluene-d8	7.320	98	338371	4.537	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.628	79	27499	3.644	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.800%
46) 2-Hexanone-d5	8.114	63	147141	45.686	ug/L	0.03
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.380%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	76852	4.791	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.619	152	122949	4.913	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028917.D
Acq On : 12 Nov 2022 08:43
Operator : SY/MD
Sample : N5602-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AG7

Quant Time: Nov 14 00:26:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 12 03:43:58 2022
Response via : Initial Calibration

