

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111122\
 Data File : VW028882.D
 Acq On : 11 Nov 2022 17:57
 Operator : SY/MD
 Sample : N5602-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF4

Quant Time: Nov 11 22:38:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 22:27:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	236315	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	218014	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	98505	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	81570	5.662	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 113.200%		
7) Chloroethane-d5	1.568	69	75318	5.947	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 119.000%		
11) 1,1-Dichloroethene-d2	2.105	63	115532	4.400	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 88.000%		
20) 2-Butanone-d5	3.886	46	156180	64.760	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 129.520%		
24) Chloroform-d	4.346	84	175060	5.918	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 118.400%		
26) 1,2-Dichloroethane-d4	5.030	65	81333	6.111	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 122.200%		
32) Benzene-d6	5.047	84	383416	5.892	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 117.800%		
36) 1,2-Dichloropropane-d6	6.066	67	108523	5.972	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 119.400%		
41) Toluene-d8	7.313	98	336741	5.749	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 115.000%		
43) trans-1,3-Dichloroprop...	7.622	79	34108	5.755	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 115.200%		
46) 2-Hexanone-d5	8.088	63	137265	54.261	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 108.520%		
56) 1,1,2,2-Tetrachloroeth...	10.210	84	67690	5.373	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 107.400%		
66) 1,2-Dichlorobenzene-d4	11.615	152	111713	6.339	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 126.800%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028882.D
Acq On : 11 Nov 2022 17:57
Operator : SY/MD
Sample : N5602-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF4

Quant Time: Nov 11 22:38:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 22:27:09 2022
Response via : Initial Calibration

