

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
 Data File : VV028883.D
 Acq On : 11 Nov 2022 18:21
 Operator : SY/MD
 Sample : N5602-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF5

Quant Time: Nov 11 22:38:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 22:27:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	197174	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	177682	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	82625	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	88558	7.367	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 147.400%#		
7) Chloroethane-d5	1.568	69	78878	7.464	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 149.200%#		
11) 1,1-Dichloroethene-d2	2.105	63	123073	5.618	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 112.400%		
20) 2-Butanone-d5	3.886	46	169171	84.072	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 168.140%#		
24) Chloroform-d	4.342	84	191406	7.756	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 155.200%#		
26) 1,2-Dichloroethane-d4	5.031	65	87937	7.919	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 158.400%#		
32) Benzene-d6	5.047	84	412935	7.786	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 155.800%#		
36) 1,2-Dichloropropane-d6	6.066	67	119057	8.039	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 160.800%#		
41) Toluene-d8	7.313	98	359544	7.531	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 150.600%#		
43) trans-1,3-Dichloroprop...	7.622	79	35796	7.411	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 148.200%#		
46) 2-Hexanone-d5	8.088	63	147532	71.557	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 143.120%#		
56) 1,1,2,2-Tetrachloroeth...	10.210	84	78976	7.692	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 153.800%#		
66) 1,2-Dichlorobenzene-d4	11.615	152	118167	7.994	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 159.800%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028883.D
Acq On : 11 Nov 2022 18:21
Operator : SY/MD
Sample : N5602-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF5

Quant Time: Nov 11 22:38:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 22:27:09 2022
Response via : Initial Calibration

