

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111122\
 Data File : VW028887.D
 Acq On : 11 Nov 2022 20:00
 Operator : SY/MD
 Sample : N5602-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF9

Quant Time: Nov 11 22:39:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 22:27:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	291115	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	264208	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	128189	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	86366	4.866	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.400%
7) Chloroethane-d5	1.564	69	79914	5.122	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.400%
11) 1,1-Dichloroethene-d2	2.105	63	119183	3.685	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.600%
20) 2-Butanone-d5	3.902	46	152641	51.379	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.760%
24) Chloroform-d	4.342	84	188172	5.164	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
26) 1,2-Dichloroethane-d4	5.031	65	86674	5.287	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
32) Benzene-d6	5.047	84	400694	5.081	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	6.066	67	117875	5.353	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.000%
41) Toluene-d8	7.313	98	358193	5.046	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.000%
43) trans-1,3-Dichloroprop...	7.622	79	35265	4.910	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.200%
46) 2-Hexanone-d5	8.088	63	162279	52.933	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.860%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	80836	5.295	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	11.616	152	129135	5.631	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111122\
Data File : VV028887.D
Acq On : 11 Nov 2022 20:00
Operator : SY/MD
Sample : N5602-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF9

Quant Time: Nov 11 22:39:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 22:27:09 2022
Response via : Initial Calibration

