

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111219\
 Data File : VV013584.D
 Acq On : 12 Nov 2019 12:29
 Operator : SY/MD
 Sample : VV1112WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK52

Quant Time: Nov 13 04:51:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 13 04:27:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	364203	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	375755	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	170466	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	72286	4.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.20%
7) Chloroethane-d5	1.58	69	81563	5.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.60%
11) 1,1-Dichloroethene-d2	2.12	63	99610	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	3.96	46	315704	57.37	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.74%
24) Chloroform-d	4.39	84	232090	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.60%
26) 1,2-Dichloroethane-d4	5.08	65	112759	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.20%
32) Benzene-d6	5.09	84	357853	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	6.11	67	129670	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.35	98	278984	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
43) trans-1,3-Dichloropropene-	7.66	79	45805	4.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.80%
46) 2-Hexanone-d5	8.13	63	224420	56.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	117801	5.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	148869	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111219\
Data File : VV013584.D
Acq On : 12 Nov 2019 12:29
Operator : SY/MD
Sample : VV1112WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK52

Quant Time: Nov 13 04:51:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 13 04:27:44 2019
Response via : Initial Calibration

