

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111219\
 Data File : VV013592.D
 Acq On : 12 Nov 2019 16:30
 Operator : SY/MD
 Sample : K5743-03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Nov 13 06:05:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 13 04:27:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.69	114	390770	5.00	ug/L	0.04
28) Chlorobenzene-d5	8.91	117	369029	5.00	ug/L	0.02
60) 1,4-Dichlorobenzene-d4	11.30	152	165421	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.34	65	66895	4.06	ug/L	0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.20%
7) Chloroethane-d5	1.60	69	71716	4.90	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.00%
11) 1,1-Dichloroethene-d2	2.15	63	90972	2.79	ug/L	0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.80%#
20) 2-Butanone-d5	4.03	46	298481	50.55	ug/L	0.09
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.10%
24) Chloroform-d	4.42	84	188458	4.41	ug/L	0.02
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
26) 1,2-Dichloroethane-d4	5.10	65	96557	4.32	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
32) Benzene-d6	5.11	84	320082	4.28	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
36) 1,2-Dichloropropane-d6	6.17	67	135511	5.38	ug/L	0.05
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.60%
41) Toluene-d8	7.40	98	266072	4.12	ug/L	0.04
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	7.69	79	46031	4.75	ug/L	0.04
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
46) 2-Hexanone-d5	8.16	63	252462	65.01	ug/L	0.03
Spiked Amount	50.000	Range	45 - 130	Recovery	=	130.02%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	120974	6.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.40%#
64) 1,2-Dichlorobenzene-d4	11.67	152	147524	5.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111219\
Data File : VV013592.D
Acq On : 12 Nov 2019 16:30
Operator : SY/MD
Sample : K5743-03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Nov 13 06:05:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 13 04:27:44 2019
Response via : Initial Calibration

