


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV111219\
 Data File : VV013582.D
 Acq On : 12 Nov 2019 08:30
 Operator : SY/MD
 Sample : BFB59
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB59

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Nov 13 16:12:11 2019



AutoFind: Scans 2806, 2807, 2808; Background Corrected with Scan 2793

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	27581	PASS
75	95	30	80	49.5	74458	PASS
95	95	100	100	100.0	150570	PASS
96	95	5	9	6.6	10000	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	97.1	146197	PASS
175	174	5	9	7.3	10642	PASS
176	174	95	101	95.3	139264	PASS
177	176	5	9	6.7	9324	PASS

m/z	Abundance
35.60	131.0
38.00	140.0
38.80	166.0
39.20	120.0
39.90	896.0
40.90	161.0
41.50	131.0
42.40	191.0
44.00	930.0
45.30	129.0
47.50	116.0
48.20	143.0
48.80	153.0
51.10	128.0
51.50	120.0
52.50	135.0
53.40	133.0
55.20	194.0
57.00	118.0
58.00	181.0
60.00	113.0
61.40	144.0
64.20	153.0
68.80	153.0
72.40	149.0
74.80	126.0
76.60	149.0
77.60	142.0
81.50	107.0
84.10	158.0
86.70	110.0
89.50	114.0
93.70	109.0
97.70	149.0
117.70	110.0
126.90	103.0
132.80	127.0
140.50	104.0
151.10	156.0
185.90	119.0
190.80	114.0
251.90	112.0
282.70	143.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59

m/z	Abundance
36.00	1453.0
37.10	7128.0
38.10	6506.0
39.10	2558.0
40.00	1152.0
41.60	156.0
42.00	135.0
42.50	167.0
43.30	269.0
44.00	1543.0
45.00	1212.0
46.20	317.0
47.10	1709.0
47.90	847.0
49.00	5526.0
50.00	26832.0
51.10	8819.0
52.20	501.0
53.20	246.0
53.90	208.0
55.00	616.0
56.00	1906.0
57.10	3592.0
58.00	270.0
59.30	148.0
60.10	1941.0
61.10	6964.0
62.00	6503.0
63.10	6112.0
64.00	1283.0
65.00	1722.0
65.90	274.0
67.00	431.0
68.00	13926.0
69.10	14134.0
70.10	1075.0
71.00	206.0
72.00	922.0
73.10	6618.0
74.00	23816.0
75.00	73416.0
76.10	7593.0
76.90	921.0
78.00	658.0
78.90	4748.0
80.00	1360.0
81.00	4959.0
82.00	997.0
83.10	222.0
84.20	110.0
85.60	116.0
87.00	6542.0
87.90	6016.0
91.00	448.0
92.00	4773.0
93.00	6415.0
94.00	18368.0
95.00	149184.0
96.00	10247.0
96.90	235.0
98.40	179.0
102.60	119.0
104.00	597.0
104.80	197.0
105.10	196.0
105.90	598.0
106.90	187.0
108.20	125.0
109.90	209.0
110.90	128.0
111.70	123.0
112.90	235.0
114.80	146.0
115.90	566.0
116.90	1080.0
117.80	600.0
119.00	923.0
119.80	100.0
121.80	130.0
123.80	241.0
127.80	485.0
128.80	272.0
130.00	706.0
130.80	301.0
131.70	100.0
134.80	464.0
137.00	286.0
138.80	184.0
139.80	130.0
141.00	1165.0
142.00	172.0
143.00	2625.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59

143.90	290.0
144.90	930.0
145.90	312.0
147.10	234.0
148.10	384.0
149.80	311.0
152.90	254.0
153.90	163.0
154.90	362.0
156.80	374.0
158.90	301.0
159.70	147.0
161.00	241.0
167.70	103.0
169.20	102.0
171.00	611.0
171.90	1956.0
174.00	140864.0
175.00	10636.0
176.00	133248.0
177.00	9031.0
178.10	383.0
209.20	151.0
211.80	102.0
218.10	117.0
219.30	132.0
252.40	100.0
262.60	133.0
281.30	200.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59

m/z	Abundance
36.00	1821.0
37.10	7089.0
38.10	6689.0
39.10	2443.0
40.00	1099.0
41.70	155.0
42.00	151.0
42.50	210.0
43.30	251.0
44.00	1508.0
45.10	1198.0
46.20	390.0
47.00	1560.0
47.90	905.0
49.00	6043.0
50.00	28464.0
51.00	9151.0
52.20	523.0
53.10	288.0
53.90	176.0
54.90	702.0
56.00	1896.0
57.10	3643.0
57.90	218.0
58.70	164.0
60.10	1792.0
61.10	7046.0
62.00	7524.0
63.10	6601.0
64.00	1328.0
65.00	1724.0
66.00	275.0
67.00	448.0
68.00	15196.0
69.10	14876.0
70.10	1128.0
70.80	176.0
72.00	972.0
73.00	6805.0
74.10	26008.0
75.00	77256.0
76.10	7710.0
77.00	815.0
78.00	558.0
78.90	4962.0
80.00	1378.0
81.00	5146.0
82.00	1031.0
83.10	265.0
84.20	125.0
85.60	151.0
86.90	6842.0
88.00	6018.0
91.00	454.0
92.00	4702.0
93.00	6621.0
94.10	18808.0
95.00	155072.0
96.00	10359.0
97.00	272.0
98.40	128.0
102.60	110.0
104.00	563.0
104.70	236.0
105.20	184.0
105.90	706.0
107.00	213.0
108.10	168.0
110.00	207.0
110.90	195.0
111.70	110.0
112.10	123.0
112.90	246.0
114.80	138.0
116.00	700.0
116.90	1013.0
117.80	541.0
119.00	979.0
119.90	103.0
121.90	138.0
123.80	250.0
127.70	487.0
128.80	286.0
130.00	822.0
130.80	312.0
133.60	101.0
134.70	453.0
137.00	276.0
138.90	147.0
139.80	164.0
141.00	1287.0
142.00	195.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59

143.00	2795.0
144.00	333.0
144.90	1139.0
145.90	299.0
147.10	241.0
148.10	517.0
148.70	101.0
149.80	368.0
152.00	110.0
152.90	259.0
153.90	211.0
154.80	322.0
155.20	304.0
155.70	173.0
156.90	404.0
159.00	301.0
159.70	140.0
161.10	256.0
170.90	620.0
171.90	1963.0
174.00	151360.0
175.00	10882.0
176.00	144192.0
177.00	9605.0
178.10	393.0
187.00	129.0
219.30	134.0
241.30	105.0
262.60	136.0
281.30	124.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59

m/z	Abundance
36.00	1854.0
37.10	6773.0
38.10	5969.0
39.20	2282.0
40.00	1114.0
41.70	120.0
42.40	211.0
43.30	195.0
44.10	1415.0
45.10	1200.0
46.20	385.0
47.00	1356.0
47.90	925.0
49.00	5624.0
50.00	27448.0
51.00	8395.0
52.10	450.0
53.00	355.0
54.20	156.0
54.90	683.0
56.00	1699.0
57.00	3329.0
57.80	193.0
58.40	207.0
58.80	189.0
60.10	1499.0
61.10	6566.0
62.00	7657.0
63.10	6450.0
64.00	1393.0
65.00	1479.0
66.10	318.0
66.90	454.0
68.00	14218.0
69.00	14073.0
70.10	1139.0
70.70	193.0
72.00	935.0
73.00	6595.0
74.10	25264.0
75.00	73080.0
76.00	6780.0
77.10	695.0
78.00	619.0
78.90	4983.0
80.00	1368.0
81.00	4917.0
81.90	983.0
83.10	277.0
85.60	214.0
86.20	299.0
87.00	6490.0
88.00	5937.0
90.80	462.0
92.10	4152.0
93.00	6371.0
94.10	17352.0
95.00	147456.0
96.00	9395.0
97.00	294.0
104.00	435.0
104.70	233.0
105.90	686.0
106.90	156.0
107.50	120.0
108.00	188.0
110.10	173.0
111.00	227.0
112.10	191.0
112.90	207.0
114.80	160.0
116.00	709.0
117.00	935.0
118.00	540.0
118.90	1045.0
123.90	194.0
127.70	508.0
128.00	488.0
128.80	250.0
130.00	736.0
130.80	322.0
134.70	421.0
137.00	278.0
138.90	107.0
139.80	179.0
141.00	1256.0
142.00	254.0
143.00	2761.0
144.00	337.0
144.80	1221.0
145.90	336.0
147.00	215.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59

148.10	529.0
148.80	120.0
149.80	346.0
152.00	140.0
152.90	253.0
153.90	210.0
154.80	341.0
155.30	347.0
155.80	175.0
156.90	380.0
159.00	243.0
161.00	239.0
170.90	634.0
171.90	1782.0
173.90	146368.0
175.00	10410.0
176.00	140352.0
177.00	9338.0
178.10	351.0
187.00	139.0
241.30	102.0
269.90	103.0
290.80	118.0

Instrument :
MSVOA_V
ClientSampleId :
BFB59