

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111220\
 Data File : VV019358.D
 Acq On : 12 Nov 2020 17:46
 Operator : SY/MD
 Sample : VIBLK98
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK98

Quant Time: Nov 13 02:13:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 13 01:06:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	207614	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	208693	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	112855	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	74761	47.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.54%
7) Chloroethane-d5	1.58	69	63601	50.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.02%
11) 1,1-Dichloroethene-d2	2.12	63	121508	37.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.46%
21) 2-Butanone-d5	3.91	46	85849	106.38	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.38%
24) Chloroform-d	4.38	84	146118	48.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.62%
26) 1,2-Dichloroethane-d4	5.06	65	104588	50.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.32%
32) Benzene-d6	5.07	84	265286	49.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.58%
36) 1,2-Dichloropropane-d6	6.09	67	76242	51.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.06%
41) Toluene-d8	7.34	98	252994	48.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.18%
43) trans-1,3-Dichloropropene-	7.64	79	44479	46.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.30%
47) 2-Hexanone-d5	8.11	63	55197	98.52	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.52%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	97281	47.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.34%
64) 1,2-Dichlorobenzene-d4	11.65	152	114408	53.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.14%

Target Compounds					Ovalue
46) Tetrachloroethene	8.00	164	1119	0.800 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111220\
Data File : VV019358.D
Acq On : 12 Nov 2020 17:46
Operator : SY/MD
Sample : VIBLK98
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK98

Quant Time: Nov 13 02:13:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 13 01:06:25 2020
Response via : Initial Calibration

