

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111220\
 Data File : VV019353.D
 Acq On : 12 Nov 2020 15:35
 Operator : SY/MD
 Sample : VV1112MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK81

Quant Time: Nov 13 01:22:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 13 01:06:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	207541	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	205216	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	113468	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	74460	47.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.20%
7) Chloroethane-d5	1.58	69	61823	48.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.26%
11) 1,1-Dichloroethene-d2	2.12	63	118887	36.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.86%
21) 2-Butanone-d5	3.91	46	79925	99.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.07%
24) Chloroform-d	4.38	84	140062	46.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.62%
26) 1,2-Dichloroethane-d4	5.06	65	99287	48.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.22%
32) Benzene-d6	5.07	84	255370	48.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.50%
36) 1,2-Dichloropropane-d6	6.09	67	72488	49.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.68%
41) Toluene-d8	7.34	98	241788	46.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.48%
43) trans-1,3-Dichloropropene-	7.64	79	43650	46.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.12%
47) 2-Hexanone-d5	8.11	63	51893	94.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.19%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	91568	45.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	113147	52.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111220\
Data File : VV019353.D
Acq On : 12 Nov 2020 15:35
Operator : SY/MD
Sample : VV1112MBL01
Misc : 5.0µL/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK81

Quant Time: Nov 13 01:22:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 13 01:06:25 2020
Response via : Initial Calibration

