

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111221\
 Data File : VV023419.D
 Acq On : 12 Nov 2021 04:59
 Operator : SY/MD
 Sample : VV1112WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK250

Quant Time: Nov 12 05:31:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 12 04:43:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	110397	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	106122	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	44063	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	31873	4.609	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.200%	
7) Chloroethane-d5	1.571	69	25811	4.579	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.600%	
11) 1,1-Dichloroethene-d2	2.111	63	44618	3.446	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.000%	
20) 2-Butanone-d5	3.982	46	58223	48.866	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.740%	
24) Chloroform-d	4.349	84	71190	4.830	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.600%	
26) 1,2-Dichloroethane-d4	5.043	65	33020	4.982	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.600%	
32) Benzene-d6	5.050	84	132803	4.877	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.600%	
36) 1,2-Dichloropropane-d6	6.082	67	39840	4.970	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.400%	
41) Toluene-d8	7.323	98	106626	4.179	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	7.635	79	12465	4.101	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.000%	
46) 2-Hexanone-d5	8.107	63	38396	34.336	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	68.680%	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	24696	4.284	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	37560	5.119	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111221\
Data File : VV023419.D
Acq On : 12 Nov 2021 04:59
Operator : SY/MD
Sample : VV1112WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK250

Quant Time: Nov 12 05:31:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 12 04:43:24 2021
Response via : Initial Calibration

