

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111221\
 Data File : VW023458.D
 Acq On : 12 Nov 2021 22:01
 Operator : SY/MD
 Sample : M4616-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFXR3

Quant Time: Nov 13 01:59:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 13 01:39:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	114622	5.000	ug/L	-0.02
28) Chlorobenzene-d5	8.853	117	114048	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	50366	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	47507	6.616	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	132.400%#
7) Chloroethane-d5	1.568	69	36305	6.204	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	124.000%
11) 1,1-Dichloroethene-d2	2.108	63	61952	4.609	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	3.902	46	46078	37.247	ug/L	-0.09
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.500%
24) Chloroform-d	4.349	84	80128	5.236	ug/L	-0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.800%
26) 1,2-Dichloroethane-d4	5.034	65	37879	5.504	ug/L	-0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.000%
32) Benzene-d6	5.053	84	160244	5.476	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.600%
36) 1,2-Dichloropropane-d6	6.072	67	46333	5.379	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.600%
41) Toluene-d8	7.317	98	137024	4.997	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.000%
43) trans-1,3-Dichloroprop...	7.625	79	15847	4.852	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.000%
46) 2-Hexanone-d5	8.091	63	43563	36.249	ug/L	-0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.500%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25818	4.168	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	47938	5.716	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.400%
Target Compounds						Qvalue
16) Methylene chloride	2.510	84	5918	0.593	ug/L	96
42) Toluene	7.400	91	4001	0.117	ug/L	92
54) o-xylene	9.548	106	1110	0.084	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111221\
Data File : VV023458.D
Acq On : 12 Nov 2021 22:01
Operator : SY/MD
Sample : M4616-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFXR3

Quant Time: Nov 13 01:59:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 13 01:39:11 2021
Response via : Initial Calibration

