

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111221\
 Data File : VW023464.D
 Acq On : 13 Nov 2021 00:25
 Operator : SY/MD
 Sample : M4616-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG1Y1

Quant Time: Nov 13 02:00:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 13 01:39:11 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards							
1) 1,4-Difluorobenzene		5.619	114	116488	5.000	ug/L	-0.02
28) Chlorobenzene-d5		8.853	117	117189	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.249	152	51219	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.307	65	49501	6.783	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	135.600%	#
7) Chloroethane-d5		1.568	69	36225	6.091	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.800%	
11) 1,1-Dichloroethene-d2		2.108	63	63937	4.680	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.600%	
20) 2-Butanone-d5		3.892	46	79060	62.884	ug/L	-0.10
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.760%	
24) Chloroform-d		4.352	84	81064	5.212	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%	
26) 1,2-Dichloroethane-d4		5.034	65	40584	5.803	ug/L	-0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.000%	
32) Benzene-d6		5.053	84	164406	5.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.400%	
36) 1,2-Dichloropropane-d6		6.069	67	47894	5.411	ug/L	-0.03
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.200%	
41) Toluene-d8		7.317	98	139532	4.952	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%	
43) trans-1,3-Dichloroprop...		7.625	79	16456	4.903	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.000%	
46) 2-Hexanone-d5		8.091	63	52011	42.119	ug/L	-0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.240%	
56) 1,1,2,2-Tetrachloroeth...		10.217	84	30741	4.829	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%	
66) 1,2-Dichlorobenzene-d4		11.625	152	52078	6.106	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	122.200%	#
Target Compounds							Qvalue
2) Dichlorodifluoromethane		1.130	85	2056	0.181	ug/L	# 83
17) Methyl tert-butyl Ether		2.767	73	43481	2.844	ug/L	98
34) Trichloroethene		5.921	95	3942	0.453	ug/L	92
47) Tetrachloroethene		7.982	164	2457	0.325	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111221\
Data File : VV023464.D
Acq On : 13 Nov 2021 00:25
Operator : SY/MD
Sample : M4616-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG1Y1

Quant Time: Nov 13 02:00:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 13 01:39:11 2021
Response via : Initial Calibration

