

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111221\
 Data File : VW023467.D
 Acq On : 13 Nov 2021 01:37
 Operator : SY/MD
 Sample : M4581-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 13 02:00:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 13 01:39:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	120358	5.000	ug/L	-0.02
28) Chlorobenzene-d5	8.854	117	119996	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	52465	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31625	4.194	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%
7) Chloroethane-d5	1.568	69	28302	4.606	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.108	63	47422	3.360	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.200%
20) 2-Butanone-d5	3.892	46	66315	51.051	ug/L	-0.10
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.100%
24) Chloroform-d	4.352	84	72678	4.523	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.400%
26) 1,2-Dichloroethane-d4	5.037	65	35156	4.865	ug/L	-0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
32) Benzene-d6	5.053	84	137555	4.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.072	67	42581	4.698	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.317	98	116010	4.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	7.625	79	13434	3.909	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.092	63	44068	34.852	ug/L	-0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.700%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25771	3.954	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	44283	5.069	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111221\
Data File : VV023467.D
Acq On : 13 Nov 2021 01:37
Operator : SY/MD
Sample : M4581-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 13 02:00:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 13 01:39:11 2021
Response via : Initial Calibration

