

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111225\
 Data File : VV039364.D
 Acq On : 12 Nov 2025 10:04
 Operator : SY/MD
 Sample : VV1112MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV1112MBL01

Quant Time: Nov 14 00:17:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V111125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 14 00:16:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.593	168	687450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.526	114	1110064	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1038812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	504387	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.934	65	482546	56.898	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	113.800%
35) Dibromofluoromethane	4.497	113	409864	48.140	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	96.280%
50) Toluene-d8	7.233	98	1336833	47.515	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	95.020%
62) 4-Bromofluorobenzene	9.998	95	511162	47.282	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	94.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111225\
Data File : VV039364.D
Acq On : 12 Nov 2025 10:04
Operator : SY/MD
Sample : VV1112MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV1112MBL01

Quant Time: Nov 14 00:17:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V111125W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 14 00:16:53 2025
Response via : Initial Calibration

