

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111225\
 Data File : VV039363.D
 Acq On : 12 Nov 2025 08:48
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 14 00:17:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V111125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 14 00:16:53 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.298	0.259	13.1	102	0.00
3 P	Chloromethane	0.279	0.241	13.6	101	0.00
4 C	Vinyl Chloride	0.344	0.301	12.5#	100	0.00
5 T	Bromomethane	0.205	0.192	6.3	104	0.00
6 T	Chloroethane	0.254	0.228	10.2	102	0.00
7 T	Trichlorofluoromethane	0.700	0.625	10.7	101	0.00
8 T	Diethyl Ether	0.267	0.244	8.6	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.407	11.5	101	0.00
10 T	Methyl Iodide	0.283	0.313	-10.6	108	0.00
11 T	Tert butyl alcohol	0.066	0.068	-3.0	105	0.00
12 CM	1,1-Dichloroethene	0.364	0.317	12.9#	101	0.00
13 T	Acrolein	0.006	0.006	0.0	125	0.00
14 T	Allyl chloride	0.649	0.613	5.5	103	0.00
15 T	Acrylonitrile	0.256	0.254	0.8	103	0.00
16 T	Acetone	0.326	0.318	2.5	123	0.00
17 T	Carbon Disulfide	0.575	0.474	17.6	100	0.00
18 T	Methyl Acetate	0.874	0.991	-13.4	105	0.00
19 T	Methyl tert-butyl Ether	1.369	1.401	-2.3	102	0.00
20 T	Methylene Chloride	0.504	0.420	16.7	101	0.00
21 T	trans-1,2-Dichloroethene	0.361	0.338	6.4	102	0.00
22 T	Diisopropyl ether	1.409	1.505	-6.8	103	0.00
23 T	Vinyl Acetate	0.995	1.003	-0.8	102	0.00
24 P	1,1-Dichloroethane	0.857	0.799	6.8	103	0.00
25 T	2-Butanone	0.341	0.356	-4.4	111	0.00
26 T	2,2-Dichloropropane	0.721	0.705	2.2	105	0.00
27 T	cis-1,2-Dichloroethene	0.490	0.470	4.1	102	0.00
28 T	Bromochloromethane	0.375	0.389	-3.7	105	0.00
29 T	Tetrahydrofuran	0.191	0.199	-4.2	105	0.00
30 C	Chloroform	0.946	0.892	5.7#	104	0.00
31 T	Cyclohexane	0.536	0.464	13.4	98	0.00
32 T	1,1,1-Trichloroethane	0.791	0.749	5.3	103	0.00
33 S	1,2-Dichloroethane-d4	0.617	0.583	5.5	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.383	0.393	-2.6	103	0.00
36 T	1,1-Dichloropropene	0.330	0.345	-4.5	101	0.00
37 T	Ethyl Acetate	0.507	0.440	13.2	97	0.00
38 T	Carbon Tetrachloride	0.491	0.481	2.0	101	0.00
39 T	Methylcyclohexane	0.352	0.346	1.7	98	0.00
40 TM	Benzene	1.155	1.176	-1.8	103	0.00
41 T	Methacrylonitrile	0.181	0.222	-22.7#	106	0.00
42 TM	1,2-Dichloroethane	0.450	0.460	-2.2	104	0.00
43 T	Isopropyl Acetate	0.676	0.723	-7.0	107	0.00
44 TM	Trichloroethene	0.313	0.304	2.9	99	0.00
45 C	1,2-Dichloropropane	0.313	0.345	-10.2#	102	0.00
46 T	Dibromomethane	0.251	0.245	2.4	103	0.00
47 T	Bromodichloromethane	0.544	0.554	-1.8	104	0.00
48 T	Methyl methacrylate	0.315	0.339	-7.6	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111225\
 Data File : VV039363.D
 Acq On : 12 Nov 2025 08:48
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 14 00:17:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V111125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 14 00:16:53 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.004	20.0	98	0.00
50 S	Toluene-d8	1.267	1.291	-1.9	103	0.00
51 T	4-Methyl-2-Pentanone	0.462	0.537	-16.2	107	0.00
52 CM	Toluene	0.724	0.770	-6.4#	103	0.00
53 T	t-1,3-Dichloropropene	0.468	0.510	-9.0	102	0.00
54 T	cis-1,3-Dichloropropene	0.497	0.554	-11.5	106	0.00
55 T	1,1,2-Trichloroethane	0.366	0.376	-2.7	107	0.00
56 T	Ethyl methacrylate	0.442	0.494	-11.8	101	0.00
57 T	1,3-Dichloropropane	0.564	0.592	-5.0	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.126	0.285	-126.2#	209#	0.00
59 T	2-Hexanone	0.330	0.408	-23.6#	111	0.00
60 T	Dibromochloromethane	0.470	0.480	-2.1	104	0.00
61 T	1,2-Dibromoethane	0.357	0.373	-4.5	105	0.00
62 S	4-Bromofluorobenzene	0.487	0.520	-6.8	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.309	0.288	6.8	100	0.00
65 PM	Chlorobenzene	0.995	0.976	1.9	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.415	0.410	1.2	105	0.00
67 C	Ethyl Benzene	1.440	1.519	-5.5#	102	0.00
68 T	m/p-Xylenes	0.544	0.615	-13.1	104	0.00
69 T	o-Xylene	0.514	0.573	-11.5	102	0.00
70 T	Styrene	0.920	1.082	-17.6	104	0.00
71 P	Bromoform	0.386	0.389	-0.8	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	2.442	2.561	-4.9	103	0.00
74 T	N-amyl acetate	0.957	1.001	-4.6	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.063	0.971	8.7	107	0.00
76 T	1,2,3-Trichloropropane	0.776	0.722	7.0	106	0.00
77 T	Bromobenzene	0.754	0.724	4.0	104	0.00
78 T	n-propylbenzene	2.797	3.022	-8.0	104	0.00
79 T	2-Chlorotoluene	1.790	1.842	-2.9	104	0.00
80 T	1,3,5-Trimethylbenzene	1.996	2.180	-9.2	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.290	0.317	-9.3	110	0.00
82 T	4-Chlorotoluene	2.002	2.176	-8.7	105	0.00
83 T	tert-Butylbenzene	2.194	2.249	-2.5	103	0.00
84 T	1,2,4-Trimethylbenzene	1.944	2.181	-12.2	105	0.00
85 T	sec-Butylbenzene	2.602	2.764	-6.2	104	0.00
86 T	p-Isopropyltoluene	2.245	2.411	-7.4	104	0.00
87 T	1,3-Dichlorobenzene	1.445	1.416	2.0	104	0.00
88 T	1,4-Dichlorobenzene	1.576	1.467	6.9	104	0.00
89 T	n-Butylbenzene	2.002	2.147	-7.2	104	0.00
90 T	Hexachloroethane	0.560	0.504	10.0	105	0.00
91 T	1,2-Dichlorobenzene	1.415	1.371	3.1	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.219	0.195	11.0	107	0.00
93 T	1,2,4-Trichlorobenzene	0.897	0.946	-5.5	104	0.00
94 T	Hexachlorobutadiene	0.479	0.439	8.4	105	0.00
95 T	Naphthalene	2.405	2.755	-14.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111225\
Data File : VV039363.D
Acq On : 12 Nov 2025 08:48
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
LabSampleId :
VSTDCCC050

Quant Time: Nov 14 00:17:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V111125W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 14 00:16:53 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.926	0.989	-6.8	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6