

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111319\
 Data File : VV013607.D
 Acq On : 13 Nov 2019 16:52
 Operator : SY/MD
 Sample : K5815-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A5460

Manual Integrations
 APPROVED

MMDadoda
 11/14/2019 11:15:04 AM

Quant Time: Nov 14 06:32:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Nov 14 06:01:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.68	114	379051	5.00	ug/L	-0.02
28) Chlorobenzene-d5	8.91	117	335283	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	132198	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.36	65	110452	4.54	ug/L	0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.80%
7) Chloroethane-d5	1.62	69	89008	4.59	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.80%
11) 1,1-Dichloroethene-d2	2.16	63	139332	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	4.04	46	243791	55.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.82%
24) Chloroform-d	4.43	84	187666	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
26) 1,2-Dichloroethane-d4	5.10	65	93121	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
32) Benzene-d6	5.11	84	366392	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	6.15	67	132477	4.78	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.60%
41) Toluene-d8	7.40	98	321495	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.60%
43) trans-1,3-Dichloropropene-	7.70	79	43399	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.16	63	187809	45.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.32%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	91108	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	122010	5.23	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.30	43	10406m	3.691	ug/L	
17) Methyl tert-butyl Ether	2.86	73	28511	0.553	ug/L #	95
68) 1,2,4-trichlorobenzene	13.31	180	3076	0.115	ug/L #	91
70) 1,2,3-Trichlorobenzene	13.79	180	2320	0.093	ug/L #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111319\
Data File : VV013607.D
Acq On : 13 Nov 2019 16:52
Operator : SY/MD
Sample : K5815-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A5460

Manual Integrations
APPROVED

MMDadoda
11/14/2019 11:15:04 AM

Quant Time: Nov 14 06:32:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 14 06:01:41 2019
Response via : Initial Calibration

