

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111320\
 Data File : VV019368.D
 Acq On : 13 Nov 2020 10:21
 Operator : SY/MD
 Sample : VV1113MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK82

Quant Time: Nov 16 15:07:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 16 13:41:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	214679	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	218344	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	116736	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	77596	47.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.90%
7) Chloroethane-d5	1.58	69	64997	49.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.86%
11) 1,1-Dichloroethene-d2	2.12	63	123057	36.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.92%
21) 2-Butanone-d5	3.91	46	80298	96.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.22%
24) Chloroform-d	4.37	84	146649	47.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.76%
26) 1,2-Dichloroethane-d4	5.06	65	103027	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.52%
32) Benzene-d6	5.07	84	261877	46.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.02%
36) 1,2-Dichloropropane-d6	6.09	67	75732	48.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.90%
41) Toluene-d8	7.34	98	248273	45.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.22%
43) trans-1,3-Dichloropropene-	7.64	79	43601	43.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.42%
47) 2-Hexanone-d5	8.11	63	49627	84.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.66%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	96060	44.99	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.98%
64) 1,2-Dichlorobenzene-d4	11.65	152	115856	51.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.92%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111320\
Data File : VV019368.D
Acq On : 13 Nov 2020 10:21
Operator : SY/MD
Sample : VV1113MBL01
Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK82

Quant Time: Nov 16 15:07:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 16 13:41:50 2020
Response via : Initial Calibration

