

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV111320\
 Data File : VV019375.D
 Acq On : 13 Nov 2020 14:08
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK82

Quant Time: Nov 16 15:39:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 16 13:41:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	210938	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	213923	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	110291	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	65573	41.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.48%
7) Chloroethane-d5	1.58	69	58347	45.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.32%
11) 1,1-Dichloroethene-d2	2.12	63	105374	32.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.42%
21) 2-Butanone-d5	3.90	46	76679	93.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.52%
24) Chloroform-d	4.37	84	137620	45.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.50%
26) 1,2-Dichloroethane-d4	5.05	65	100143	47.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.48%
32) Benzene-d6	5.07	84	247502	44.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.72%
36) 1,2-Dichloropropane-d6	6.09	67	72748	47.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.00%
41) Toluene-d8	7.33	98	230826	42.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.62%
43) trans-1,3-Dichloropropene-	7.64	79	41693	42.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.32%
47) 2-Hexanone-d5	8.11	63	48266	84.04	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.04%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	89970	43.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.02%
64) 1,2-Dichlorobenzene-d4	11.64	152	107389	50.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.94%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV111320\
Data File : VV019375.D
Acq On : 13 Nov 2020 14:08
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK82

Quant Time: Nov 16 15:39:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM111220WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 16 13:41:50 2020
Response via : Initial Calibration

