

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111323\
 Data File : VV032930.D
 Acq On : 13 Nov 2023 14:41
 Operator : SY/MD
 Sample : 05259-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 14 00:18:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 13 23:16:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	426879	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	429601	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	206911	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	155271	43.007	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.020%	
7) Chloroethane-d5	1.532	69	127780	42.716	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	85.440%	
11) 1,1-Dichloroethene-d2	2.060	65	67636	42.997	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	86.000%	
21) 2-Butanone-d5	3.783	46	122187	78.954	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	78.950%	
24) Chloroform-d	4.249	84	263019	42.169	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.340%	
26) 1,2-Dichloroethane-d4	4.941	65	161094	42.447	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.900%	
32) Benzene-d6	4.960	84	515786	44.021	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.040%	
36) 1,2-Dichloropropane-d6	5.989	67	152023	44.637	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	89.280%	
41) Toluene-d8	7.243	98	460365	42.616	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	85.240%	
43) trans-1,3-Dichloroprop...	7.551	79	74451	41.177	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	82.360%	
47) 2-Hexanone-d5	8.024	63	106554	85.399	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	85.400%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	204586	41.007	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	82.020%	
66) 1,2-Dichlorobenzene-d4	11.561	152	187443	46.622	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111323\
Data File : VV032930.D
Acq On : 13 Nov 2023 14:41
Operator : SY/MD
Sample : 05259-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 14 00:18:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 13 23:16:30 2023
Response via : Initial Calibration

